Downloaded on 30 September 2024 09:28 | Tri Okta Ervina - ENG BJS

Parts List

SICK_F01_003

Device tag	Placement	Quantity QU	Designation	SAP number	Type number Order number	Manufacturer
BQ1	+ORT/103.1	1 Pieces	MCS100E		MCS100E	SICK AG
EA1	+ORT/102.7	1 Pieces	Enclosure light 18W with motion detector and power socket	6029407	SZ 4155.110 4155110	Rittal
EB42	+ORT/119.1	1 Pieces	Gas sampling filter 230V 2m	1042984		SICK AG
EB51	+ORT/121.1	1 Pieces	sample gas line 230V 2HC			Eltherm Elektrowärmetechnik
FC1	+ORT/101.1	1 Pieces	Circuit breaker 1~, B 6A, 6kA	6029689	S201-B6 2CDS251001R0065	ABB Stotz-Kontakt
FC01	+ORT/100.3	1 Pieces	Circuit breaker 1~, B 6A, 6kA	6029689	S201-B6 2CDS251001R0065	ABB Stotz-Kontakt
FC2	+ORT/101.4	1 Pieces	Circuit breaker 1~, B 10A, 6kA	6029690	S201-B10 2CDS251001R0105	ABB Stotz-Kontakt
FC3	+ORT/101.6	1 Pieces	Circuit breaker 1~, B 6A, 6kA	6029689	S201-B6 2CDS251001R0065	ABB Stotz-Kontakt
FC10	+ORT/102.1	1 Pieces	Circuit breaker 3~, B 20A, 6kA	6029703	S203-B20 2CDS253001R0205	ABB Stotz-Kontakt
FC11	+ORT/102.3	1 Pieces	Circuit breaker 1~, B 10A, 6kA	6029690	S201-B10 2CDS251001R0105	ABB Stotz-Kontakt
FC12	+ORT/102.5	1 Pieces	Circuit breaker 1~, B 6A, 6kA	6029689	S201-B6 2CDS251001R0065	ABB Stotz-Kontakt
FC13	+ORT/102.7	1 Pieces	Circuit breaker 1~, B 10A, 6kA	6029690	S201-B10 2CDS251001R0105	ABB Stotz-Kontakt
GQ1	+ORT/118.1	1 Pieces	Diaphragm vacuum pump, heated, max. 18,5 l/min., 1 inlet	6035824	PJ22242-N024.11 114960/000000	KNF Neuberger
GQ1	+ORT/118.1	1 Pieces	Insulation cover for pump head Ø 94x80	5313568	5313568	SICK AG
GQ3	+ORT/102.5	1 Pieces	Filter fan 230m³/h, 230V, 50/60Hz	6048627	SK.3241100 3241100	Rittal
KF1	+ORT/104.0	1 Pieces	I/O-module box	2025018	2025018	SICK AG
KF1.2	+ORT/104.2	1 Pieces	I/O-module box, digital output board with 4 potential-free change-over contacts	2025022	DOM4 2025022	SICK AG
KF1.3	+ORT/104.2	1 Pieces	I/O-module box, digital output board with 8 NO contact contacts	2025026	DOS8 2025026	SICK AG

CONFIDENTIAL

Uncontrolled when printed or downloaded. Verify that this is the correct version before use.
Reproduction of part or all the contents of the documents is prohibited without written permission from PT. Bhumi Jepara Service